



## Label Antibodies

**(VXI-7601, VXI-7604, VXI-7605, VX-6301)**

For more detailed product information and full VaxArray Influenza Pandemic Hemagglutinin potency assay procedure, please refer to the most recent revision of the VaxArray Influenza Pandemic Hemagglutinin Potency Assay Operation Manual and the VaxArray Influenza Pandemic Hemagglutinin Label Selection Guide.

### INTENDED USE

Label antibodies are for use with the VaxArray Influenza Pandemic Hemagglutinin potency assay intended for **RESEARCH USE ONLY**.

### REAGENTS & MATERIALS

- **A&B pAb Label (10X)** (VXI-7601) – A fluorescent labeled polyclonal antibody applied during label incubation. Recommended for use with A/H5, A/H7, and A/H9 split viruses and VLPs. Sufficient quantity for 32 arrays (2 slides). **Store at 2-8°C.**
- **A Hemagglutinin mAb Label (10X)** (VXI-7604) – A fluorescent labeled monoclonal antibody applied during label incubation. Recommended for use with A/H5 and A/H9 recombinant HA. Sufficient quantity for 32 arrays (2 slides). **Store at 2-8°C.**
- **A H7 mAb Label (10X)** (VXI-7605) – A fluorescent labeled monoclonal antibody applied during label incubation. Recommended for use with A/H7 recombinant HA. Sufficient quantity for 32 arrays (2 slides). **Store at 2-8°C.**
- **Fiducial Label (10X)** (VX-6301) – A fluorescent labeled antibody applied during label incubation to serve as an assay control. Sufficient quantity for 32 arrays (2 slides). **Store at 2-8°C.**

### SAFETY AND WARNINGS

Handle all biological samples according to appropriate biosafety level precautions. Biosafety guidelines can be found in the CDC's Biosafety in Microbiological and Biomedical Laboratories (BMBL) manual (<http://www.cdc.gov/biosafety/publications/bmbl5/index.htm>). Treat and dispose

of waste as necessary according to local regulations. MSDSs for individual reagents are available at <http://www.indevr.com>

### ASSAY PROCEDURE

For full Assay Procedure, please refer to the most recent revision of the full VaxArray Influenza Pandemic Hemagglutinin Potency Assay Operation Manual.

### SUPPORT INFORMATION

Detailed information regarding the VaxArray Influenza Pandemic Hemagglutinin potency assay can be found in the most recent revision of the VaxArray Influenza Pandemic Hemagglutinin Potency Assay Operation Manual and the VaxArray Influenza Pandemic Hemagglutinin Label Selection Guide. Additional information regarding performance, technical specifications, specific applications, and more can be found at [www.indevr.com](http://www.indevr.com).

For technical support please contact us at [techsupport@indevr.com](mailto:techsupport@indevr.com) or 1-303-402-9100. For sales or quote requests, please contact us at [sales@indevr.com](mailto:sales@indevr.com) or 1-303-402-9100.

Copyright 2014-2017, InDevR Inc. All rights reserved.  
For Laboratory Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.  
October 2017